

Acme Scale Company  
1801 Adams Ave  
Phone 510-638-5040

Reviewed By: \_\_\_\_\_  
Date Reviewed: \_\_\_\_\_ 365  
Inspection Intervals by Days: \_\_\_\_\_  
Date Service Performed: 9-1-10  
Calib. Performed By: A. TORRES  
Next Calib. Due By: 9-1-11

# CALIBRATION REPORT

Acme No: 16089 Make: ICS  
Building: ICSCCS5K  
Room: 1206861

Department: \_\_\_\_\_  
Asset ID: \_\_\_\_\_  
Capacity: 5000 lb x 1 lb  
MMS Cap: 5000 lb x 1 lb

Acme Scale Test Procedure No: 02  
Standards Ratio: 4:1  Yes  No  
Accuracy: +/- .1% of Reading

## Compression

Nominal Value lb	Tolerance lb	AS Found lb	AS Found +/- lb	After Adjust (As Left) lb	After Adjustment Error +/- lb
0.00	0.000				
1000.00	1.000				
2000.00	2.000				
3000.00	3.000				
4000.00	4.000				
5000.00	5.000				

## Tension

Nominal Value lb	Tolerance lb	AS Found lb	AS Found Error +/- lb	After Adjustment (As Left) lb	After Adjustment Error +/- lb
0.00	0.000	0	0	N/A	N/A
1000.00	1.000	999	-1		
2000.00	2.000	1999	-1		
3000.00	3.000	3000	0		
4000.00	4.000	4002	+2		
5000.00	5.000	5005	+5		

As Found Compression  
In Tolerance   
Out of Tolerance

As Left Compressed  
Calibrated   
With in Spec

As Found Tension  
In Tolerance   
Out of Tolerance

As Left Tension  
Calibrated   
With in Spec

Do Not Use

Notes: \_\_\_\_\_

## Standard ID

2508	16767	K1001	K2501	H10	KG-10	KG-20	KG-40	KG-50
LB10	7282	19134	20122	20129	20133	500-1	2010	2011
2012	10021	2503	2504	2507	2509	2510	2512	5041-5060
5071 - 5090	<input checked="" type="checkbox"/>	5021-5040	50120-50139	A2001-A2050	KG-70	5001-5020	C03-C04	C06, C10
20126	20127	A13-A16	2513	F696-F699	2506	TE148A-D	17194	C01, C05
C02, C09	C08, C17	KG-60	LB-20	50160-50169	C20 - C21	50150 - 50159	5061-5070	M9130
K1002								



Original

TRACEABLE CERTIFICATE OF CALIBRATION

Acme Scale Company  
1801 Adams Ave  
Phone 510-638-5040

Date: 9/1/2010

THIS IS TO CERTIFY THAT THE FOLLOWING INSTRUMENT WAS TESTED USING EQUIPMENT PERIODICALLY CALIBRATED USING WEIGHTS WITH ACCURACY TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY, AND IS GUARANTEED ACCURATE TO THE TOLERANCE INDICATED ON THE DATE TESTED IN ACCORDANCE WITH ANSI/NCSL Z540-1. THIS DOCUMENT SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN APPROVAL OF AN ACME SCALE COMPANY OFFICIAL.

Performed for:

Certificate # 65917

Accuracy: .1%

Slackwater\_SF  
Bay Area Multihull Association (BAMA)

Instrument:

Make ICS  
Model ICSCCS5K  
Capacity 5000 lb x 1 lb

Serial Number: 1206861

Asset ID

Test Weight	Test Number	Test Number	N.I.S.T. Exp. Date
2001-2009	09270		12/22/2011

As Found: In Tolerance  
As Left: In Tolerance  
Adjusted: No  
Next Due Date: 9/1/2011

Procedure: 02  
Calibrated by: Adam Tordoff  
Calibration date: 9/1/2010

Reviewed by: 

Name:  
Title:

Ron Widgren  
Service Manager

INVOICE: 002136  
SEP 01, 10 08:42  
RRN: 024416404386  
AUTH NO: 78036P  
AUS: MATCH  
CUV2 RESPONSE: MATCH

TOTAL \$240.00

THANK YOU