

Bay City Scale Calibration Report
25352A Cypress Ave, Hayward, CA 94544 510-266-0607

ISO 9001:2008

SCALE MODEL ICS CCS 5K	INDICATOR SERIAL# 1206861	DATE <i>March 19th, 2013</i>
CAPACITY/RESOLUTION 5000 X 1 LB	BASE SERIAL# 1206861	TYPE S TYPE CELL 5K
CUSTOMER Slackwater_SF	STREET ADDRESS	
CITY	STATE CA	ZIP

TESTING Check Boxes to signify work performed MINOR ADJUSTMENTS

<input checked="" type="checkbox"/> Set scale to zero	<input checked="" type="checkbox"/> Span Test	<input type="checkbox"/> Zero	<input type="checkbox"/> Span	<input checked="" type="checkbox"/> Sign-Off and date Calibration Sticker
<input type="checkbox"/> Corner Test	<input type="checkbox"/> Tare Test	<input type="checkbox"/> Corner	<input type="checkbox"/> Tare	

**TEST RESULTS
CORNER TEST**

CORNER POSITIONS

1	2
3	4

		AS FOUND		AFTER ADJUSTMENTS	
POSITION	WEIGHT*	INDICATION	ERROR	INDICATION	
1					
2					
3					
4					

SPAN TEST

		AS FOUND		AFTER ADJUSTMENTS	
WEIGHT*	INDICATION	ERROR	INDICATION		
1000 LBS	<i>1000</i>	<i>∅</i>	_____		
3000 LBS	<i>3000</i>	<i>∅</i>	_____		
5000 LBS	<i>5000</i>	<i>∅</i>	_____		

*Equals cumulative weight total.

Scale calibrated to tolerances specified by United States Department of Commerce, Technology Administration, National Institute of Standards and Technology, Handbook 44, Section 2.20, NIST

TEST WEIGHT *1000 lbs* TEST WT CERT. # CA 08017 AVOIRDUPOIS

TEST WEIGHT _____ TEST WT CERT. # *1443* - NV01001213 AVOIRDUPOIS

TEST WEIGHT _____ TEST WT CERT. # *R948* - NV01001212 AVOIRDUPOIS

TEST WEIGHT _____ TEST WT CERT. # *OMM6* - NV01001141 AVOIRDUPOIS

NEXT CALIBRATION SCHEDULED FOR *March 2014*

[Signature]
TECHNICIAN SIGNATURE ID# *07298 10638*

CUSTOMER SIGNATURE *[Signature]*
FORM 4.10.1.1.B-1 10/10